



# 中国国家强制性产品认证证书



证书编号：2021322304004376

认证委托人名称： BETA B.V.  
 认证委托人地址： Verrijn Stuartlaan 22, 2288EL Rijswijk, The Netherlands  
 生产者名称： BETA B.V.  
 生产者地址： Verrijn Stuartlaan 22, 2288EL Rijswijk, The Netherlands  
 生产企业名称： BETA B.V.  
 企业地址： Verrijn Stuartlaan 22, 2288EL Rijswijk, The Netherlands

产品名称： 压力/差压/温度开关  
 系列、规格、型号： W 系列  
 标 准： GB 3836.1-2010、GB 3836.2-2010、GB 12476.1-2013、  
 GB 12476.5-2013

上述产品符合强制性产品认证实施规则 CNCA-C23-01:2019 的要求，特发此证。

发证日期：2021 年 12 月 3 日                      有效期至：2026 年 12 月 2 日

首次发证日期：2021 年 12 月 3 日

证书有效期内本证书的有效性依据发证机构的定期监督获得保持。

本证书的相关信息可通过国家认监委网站 [www.cnca.gov.cn](http://www.cnca.gov.cn) 查询



批 准：



## 上海仪器仪表自控系统检验测试所有限公司





# CERTIFICATE FOR CHINA COMPULSORY PRODUCT CERTIFICATION



**CERTIFICATE NO: 2021322304004376**

**APPLICANT:** BETA B.V.  
**ADDRESS:** Verrijn Stuartlaan 22, 2288EL Rijswijk, The Netherlands  
**MANUFACTURER:** BETA B.V.  
**ADDRESS:** Verrijn Stuartlaan 22, 2288EL Rijswijk, The Netherlands  
**FACTORY:** BETA B.V.  
**ADDRESS:** Verrijn Stuartlaan 22, 2288EL Rijswijk, The Netherlands  
**PRODUCTNAME:** Pressure, Differential Pressure and Temperature Switches  
**SERIES,SPECIFICATION,MODEL:** W sereis  
**STANDARDS:** GB 3836.1-2010、GB 3836.2-2010、GB 12476.1-2013、  
 GB 12476.5-2013

This is to certify that the above mentioned product(s)complies with the requirements of implementation rules for compulsory certification (REFNO. CNCA-C23-01:2019).

**Valid from:**December 3, 2021

**Valid until:** December 2, 2026

**Date of original certification:**December 3, 2021

**The validity of this certificate is subject to positive result of the regular follow up inspection by issuing certification body until the expiry date.**

This certificate is available through CNCA's website: [www.cnca.gov.cn](http://www.cnca.gov.cn)



APPROVAL:

Guo AiHua



**Shanghai Inspection and Testing Institute of Instruments and Automation Systems Co., Ltd.**

<http://www.sitiias.com.cn>

Building 9, 103 Cao Bao Road, Shanghai 200233, China

Tel: +86 21 64510844

**S 0005423**





# 中国国家强制性产品认证证书



证书编号：2021322304004376

## 附件

产品名称： 压力/差压/温度开关

型号规格：

W a b c d e f g

a: 代表外壳，可选代码为 3, 8, 9;

b: 代表温度范围或压力范围，

可选代码(压力开关)：

301L/302L/304L/306L/308L/402M/406M/408M/502H/506H/508H/706H/708H/808H/908H/909H/904F/906F/908F/918F;

或可选代码(差压开关)：608M/352H/354H/356H/359H/356D/358D;

或可选代码(温度开关)：T528H/T548H/T568H/T588H;

c: 对于压力开关，代表过程连接；对于温度开关，代表传感器测温型式；

d: 代表膜片和 O 型圈；

e: 代表开关元件；

f: 代表可选项；

g: 代表与防爆无关的其它特殊要求。

防爆标志： Ex d IIC T5/T6 Gb, Ex tD A21 IP6X T100°C

电气参数： 480Va. c. 10A 50/60Hz; 125Vd. c. max. 5A

批 准：



## 上海仪器仪表自控系统检验测试所有限公司